



## National Accreditation Board for Testing and Calibration Laboratories

(A Constituent Board of Quality Council of India)



NABL / C-1245

29.03.2019

**CHANDRA SHEKHAR PRASAD**  
GAT CALIBRATION LAB (GMR AEROTECHNIC LTD)  
PLOT NO 1, GMR HYDERABAD AVIATION  
SEZ LTD  
HYDERABAD, TELANGANA-500409  
Mobile: 9910311741  
E-mail: chandrashekhar.prasad@gmraerotech.in

Subject: **Grant of Accreditation**

Dear Sir/Ma'am,

NABL has reviewed the assessment report of your laboratory and pleased to inform you the grant of accreditation in accordance with ISO/IEC 17025:2005 in the discipline of Mechanical calibration as per the revised scope and authorized signatories recommended by the assessment team. However, you are required to submit/clarify following within 30 days:

- Range of Dead Weight Tester to be segregated as per available traceability i.e. 5 psi to 1500 psi and 600 psi to 15000 psi.

Also, maximum range of Torque wrench to be given as 990 Nm with CMC 2.0%.

You may issue Calibration Certificate vide Certificate no. CC-2974 with NABL symbol w.e.f 28.03.2019 and valid till 27.03.2021.

The accreditation is granted for two years subject to your satisfactory compliance to the terms and conditions to maintaining NABL accreditation (refer NABL 131). You are required to follow NABL 133 for the using NABL symbol.

There will be an on-site surveillance visit on completion of a year of accreditation to verify your continued compliance to NABL requirements.

You are required to submit the 4 year PT/ILC plan as required by NABL -163.

Yours faithfully,  
**Rajeshwar Kumar**  
rajeshwar@nabl.qcin.org